

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

MAY 13 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130119</u>
Name <u>Tobin P Brennan</u>	ID# <u>4542</u>	Date <u>5-11-09</u>
A Agency <u>Soldotna Police Department</u>		Phone # <u>(907) 262-4455</u>
Instrument Location <u>Soldotna Police Department (44510 Sterling Hwy, Soldotna Alaska 99669)</u>		
B Alco S/N <u>X301556</u> Target Value <u>.082</u> High Pressure <u>700</u>		
Alco Test Values 1 st Alco <u>.086</u> 2 nd Alco <u>.085</u>		
Signature <u>[Signature]</u> CGB 5/18/09		
(OVER)		


I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 22 day of May, 2009.

[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130119

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130119

MAY 11, 2009

OPERATOR'S NAME:

BRENNAN

OPERATOR'S NUMBER: 4542

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/NI :

UOC

O.L. #: UOC

DEPT/AGENCY: SL51

CASE/REPORT: UOC

TEST TYPE: U

ALCO TARGET VALUE: .002

ALCO S/N: X301556

— BREATH ANALYSIS —

.002 ADJUSTED FOR 30.17 in
ALCO TARGET .002 06:04
BLANK TEST .000 06:05
INTERNAL STANDARD VERIFIED 06:05
ALCO TV 30.17 in .006 06:05
BLANK TEST .000 06:06
SUBJECT SAMPLE .000 06:07
BLANK TEST .000 06:08
ALCO TV 30.17 in .005 06:08
BLANK TEST .000 06:09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130119

MAY 11, 2009

TIME 06:02

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 03/29/00
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 44c
BAROMETER: 30.17 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!""%&'()*+,-./0123456789:;<=>?@ABCDEFGHI
HIJKL/MNOPQRSTUVWXYZ[\]^_`abdefghijklmno
pqrstuvwxyz{|}~